

3M™ Petrifilm™ Aqua Yeast and Mold Count Plates

Product Information

MEDIUM:	3M Petrifilm Aqua Yeast and Mold Count Plates (6408/6413)																
ISO/BSI CERTIFICATE OF REGISTRATION NUMBER:	Manufactured at Brookings, SD, USA (ISO 9001:2008, FM 14552)																
DATE OF EXPIRATION/ LOT NUMBER	Expiry and lot number indicated on each package. Lot number indicated on each plate.																
FORMULATION:	Nutrients supplemented with antibiotics, a cold-water soluble gel, and indicator.																
METHOD OF PREPARATION:	Nutrients and gels coated onto film. For use with membrane filtration, place filtered sample onto plate and then hydrate with one mL. See product instructions for instructions.																
CONTAMINATION CHECK:	96 randomized plates per batch tested Incubated at 25°C for 5 days Utilizes statistical sampling plan																
EFFICACY CHECK:	Organisms tested include: <table><thead><tr><th>Organism</th><th>Acceptable Batch</th></tr></thead><tbody><tr><td><i>Enterococcus faecalis</i></td><td>No Growth</td></tr><tr><td><i>Acinetobacter</i> species</td><td>No Growth</td></tr><tr><td><i>Aspergillus niger</i> M6</td><td>Productivity Ratio ≤0.7</td></tr><tr><td><i>Cladosporium cladosporioides</i> ATCC 16022</td><td>Productivity Ratio ≤0.7</td></tr><tr><td><i>Hansenula anomala</i> Y28</td><td>Productivity Ratio ≤0.7</td></tr><tr><td><i>Saccharomyces cerevisiae</i> Y41</td><td>Productivity Ratio ≤0.7</td></tr><tr><td><i>Candida albicans</i> ATCC 10231</td><td>Productivity Ratio ≤0.7</td></tr></tbody></table>	Organism	Acceptable Batch	<i>Enterococcus faecalis</i>	No Growth	<i>Acinetobacter</i> species	No Growth	<i>Aspergillus niger</i> M6	Productivity Ratio ≤0.7	<i>Cladosporium cladosporioides</i> ATCC 16022	Productivity Ratio ≤0.7	<i>Hansenula anomala</i> Y28	Productivity Ratio ≤0.7	<i>Saccharomyces cerevisiae</i> Y41	Productivity Ratio ≤0.7	<i>Candida albicans</i> ATCC 10231	Productivity Ratio ≤0.7
Organism	Acceptable Batch																
<i>Enterococcus faecalis</i>	No Growth																
<i>Acinetobacter</i> species	No Growth																
<i>Aspergillus niger</i> M6	Productivity Ratio ≤0.7																
<i>Cladosporium cladosporioides</i> ATCC 16022	Productivity Ratio ≤0.7																
<i>Hansenula anomala</i> Y28	Productivity Ratio ≤0.7																
<i>Saccharomyces cerevisiae</i> Y41	Productivity Ratio ≤0.7																
<i>Candida albicans</i> ATCC 10231	Productivity Ratio ≤0.7																
PACKAGING:	Pack size: 50 plates per foil pouch Film grade: Plastic foil laminate Seal integrity check: Pressure Decay test performed																
MEASUREMENT SYSTEMS CALIBRATION AND TRACEABILITY:	Incubator temperature 3M internal calibration Minimum calibration once per year for all equipment																
SHELF LIFE:	18 months from date of manufacture																
STORAGE CONDITIONS:	Store at temperatures less than or equal to 8°C																